



Aalborg Universitet

AALBORG UNIVERSITY
DENMARK

A Patient-Specific Musculoskeletal Model of Total Knee Arthroplasty to Predict In Vivo Knee Biomechanics

Marra, Marco; Vanheule, Valentine; Fluit, René; Koopman, Bart (H.F.J.M); Rasmussen, John; Verdonschot, Nico; Andersen, Michael Skipper

Creative Commons License
Unspecified

Publication date:
2015

Document Version
Accepted author manuscript, peer reviewed version

[Link to publication from Aalborg University](#)

Citation for published version (APA):

Marra, M., Vanheule, V., Fluit, R., Koopman, B., Rasmussen, J., Verdonschot, N., & Andersen, M. S. (2015). A Patient-Specific Musculoskeletal Model of Total Knee Arthroplasty to Predict In Vivo Knee Biomechanics. Abstract from 5th Dutch Bio-Medical Engineering Conference 2015, Egmond aan Zee, Netherlands.

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal -

Take down policy

If you believe that this document breaches copyright please contact us at vbn@aub.aau.dk providing details, and we will remove access to the work immediately and investigate your claim.

A PATIENT-SPECIFIC MUSCULOSKELETAL MODEL OF TOTAL KNEE ARTHROPLASTY TO PREDICT *IN VIVO* KNEE BIOMECHANICS

**Marco A. Marra^{*1}, Valentine Vanheule², René Fluit³, Bart H.F.J.M. Koopman³,
John Rasmussen⁴, Nico Verdonschot¹, Michael S. Andersen⁴**

¹Orthopaedic Research Laboratory, Radboudumc,
P.O. Box 9101, 6500 HB Nijmegen, The Netherlands
*e-mail: Marco.Marra@radboudumc.nl

²Materialise N.V., Technologielaan 15, 3000 Leuven, Belgium

³University of Twente, P.B. 217, Gebouw Horstring, 7500 AE Enschede, The Netherlands

⁴Aalborg University, Fibigerstrade 16, DK-9220 Aalborg East, Denmark

ABSTRACT

Musculoskeletal (MS) models are useful to gain information on *in vivo* biomechanics that would be otherwise very difficult to obtain. However, before entering the clinical routine MS models must be thoroughly validated. This study presents a novel MS modelling framework capable of integrating the patient-specific MS architecture in a very detailed way, and simultaneously simulating body level dynamics and secondary knee kinematics. The model predictions were further validated against publicly available *in vivo* experimental data.

The bone geometries were segmented from CT images of a patient with an instrumented Total Knee Arthroplasty (TKA) from the “Grand Challenge Competition to Predict *In Vivo* Knee Loads” dataset. These were inputted into an advanced morphing technique in order to scale the MS architecture of the new TLEM 2.0 model¹ to the specific patient. A detailed 11-DOF model of the knee joint was constructed that included ligaments and rigid contact. An inverse kinematic and a force-dependent kinematic technique² were utilized to simulate one gait cycle and one right-turn trial. Tibiofemoral (TF) joint contact force predictions were evaluated against experimental TF forces recorded by the TKA prosthesis, and secondary knee kinematics against experimental fluoroscopy data.

The coefficient of determination and the root-mean-square error between predicted and experimental tibiofemoral forces were larger than 0.9 and smaller than 0.3 body-weights, respectively, for both gait and right-turn trials. Secondary knee kinematics were estimated with an average Sprague and Geers’ combined error as small as 0.06.

The modelling strategy proposed permits a high level of patient-specific personalization and does not require any non-physiological parameter tuning. The very good agreement between predictions and experimental *in vivo* data is promising for the future introduction of the model into clinical applications.

REFERENCES

- [1] V. Carbone, R. Fluit, P. Pellikaan, M. M. van der Krogt, D. Janssen, M. Damsgaard, L. Vigneron, T. Feilkas, H. F. J. M. Koopman, and N. Verdonschot, “TLEM 2.0 – A comprehensive musculoskeletal geometry dataset for subject-specific modeling of lower extremity” (Manuscript submitted for publication, 2014).
- [2] M. S. Andersen, M. Damsgaard, and J. Rasmussen, “Force-dependent kinematics: a new analysis method for non-conforming joints”, XIII International Symposium on Computer Simulation in Biomechanics, 2011.